

Calibration Certificate

TEST REPORT

REPORT NO: TL20250808

MODEL: ATL302

TEST CONDITIONS:

TEMP. 25 °C

HUMIDITY: 60 %

DATE: 2025/10/28

1. APPEARANCE: PASS

2. Parameter Calibration:

Standard	Standard Point	Test resulting	error	Standard Point	Test resulting	error
□ Standard wind speed meter	□ 5m/s	4.5m/s	1.6%	□ 10m/s	10.6m/s	2%
□ Standard Humidity	□ 50%	52%	2%	□ 75%	72%	3%
□ Standard Temperature	□ 25°C	25°C	0	□ 40°C	39.8°C	0.2%
□ Standard Barometer	□ 101325Pa	101200Pa	0.2%			



3. CONCLUSION: PASS

TESTED BY 黎新安

INSPECTED BY 谢冰

APPROVED BY 石高